PHYSIO CONTROL

LIFEPAK CR® Plus and LIFEPAK EXPRESS® Defibrillators

User's Checklist

Year: Unit Serial Number: Location:

MONTH																								
DATE																								
INITIALS																								
Instruction and Recommended Corrective Action																								
Check defibrillator for damage or foreign substances.																								
OK is visible on readiness display. If not, refer to Troubleshooting Table.																								
Note use by date of electrode pads: _____																								
Extra set of electrode pads stored with AED (optional). Note use by date of electrode pads: _____																								
Infant/Child electrode pads* stored with AED (optional). Use by date: _____																								
Resuscitation Kit containing: disposable gloves, face mask, trauma scissors and razor stored with AED.																								
Other resuscitation equipment stored with AED: _____ _____																								

* The LIFEPAK CR Plus AED may be used on children up to 8 years old or 25kg (55 lbs) only with the Infant/Child Reduced Energy Defibrillation Electrodes.

Troubleshooting Readiness Display	
CHARGE-PAK™ charger and electrode pads need to be replaced OR CHARGE-PAK charger is not installed.	Use AED if needed.
Service needed.	Attempt to use AED in an emergency. If AED is inoperable, retrieve another AED if possible and continue CPR until EMS arrives.
Internal battery not fully charged.	Verify CHARGE-PAK charger is installed. If CHARGE-PAK charger installed use unit if needed.

NOTE: If corrective action is needed leave box unchecked, complete troubleshooting log, and notify: _____.

Troubleshooting Log

Date/Time	Action Taken

Signature Log

Initials	Print Full Name

Initials	Print Full Name

©2009 Physio-Control Inc. Redmond, WA 98052, USA. All rights reserved.
GDR 3302426_B